

Operations

- Front-end
 - One power supply (LBA22) is down and cannot be replaced until the shutdown.
 - Few known bad cases of corrupted data (EBC63,...)
- Back-end
 - Several DSP updates. Current code 14.6.2
 - Staging firmware updated.
- Operation
 - Standard calibration (CIS mono, CIS scan, pedestals) and laser runs every week.
 - Special CIS mono (ramp) runs for OFV. Working on defining a standard CIS ramp run
 - High rate tests with CTP+L1Calo+LAr+Tile (very helpful)
- Maintenance
 - Lab 175@CERN is up and running. Cs scans, Cosmics, DAQ development
 - TDAQ release updated to tdaq-02-00-03. Few Tile releases.